



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 8:22 am, Oct 31, 2013

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN
204004

NAME OF AGENCY
Washington Police Department

DATE OF INSPECTION
10/29/2013

LOCATION OF INSTRUMENT (STREET AND CITY)
301 Jefferson St; Washington, MO

TIME OF INSPECTION
6:57 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ **DIAGNOSTIC CHECK (PRINTOUT ATTACHED)** DATE AND TIME (from printout) 10/29/2013 / 18:57

☒ **COMPUTER**

☒ **DETECTOR**

☒ **PROGRAM**

☒ **FILTERS**

☒ **HEATERS SAMPLE CHAMBER** 50 °C

☒ **QUARTZ STANDARD**

☒ **FLOW DETECTOR**

☒ **CALIBRATION**

☒ **PUMP HIGH SPEED**

☒ **PRINTER**

☒ **INDICATOR LIGHTS**

☒ **SIMULATOR SOLUTION SUPPLIER** Repco LOT # 13002 EXP. DATE 06/19/2015

☒ **SIMULATOR TEMP** (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2232 EXP. DATE _____

☒ **CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ **0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE**

☐ **0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE**

☐ **0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE**

TEST 1 0.095

TEST 2 0.095

TEST 3 0.095

☒ **PERFORM R.F.I. TEST (PRINTOUT ATTACHED)**

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	2	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE

Sgt. Paul R. Kesterson

PRINT FULL NAME

Sgt. Paul R. Kesterson DSN 279

TYPE II PERMIT NUMBER/EXPIRATION DATE

220348 10/10/2014

TELEPHONE NUMBER

(636) 390-1063

RETURN COMPLETED REPORT TO THE:

Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
WASHINGTON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 204004
10/29/13
18:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 50c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~

Operator Signature

[Handwritten Signature] 229

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
WASHINGTON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 204004
10/29/13

TESTING OFFICER:
KESTERSON/PAUL/R
OFFICER I.D.: 279
PERMIT NUMBER: 220348
EXPIRATION DATE: 10/10/14
MISCELLANEOUS DATA:

---- SUPERVISOR MODE ----

BLANK TEST	.000	19:00
INTERNAL STANDARD	VERIFIED	19:00
EXTERNAL STANDARD	.095	19:01
BLANK TEST	.000	19:01
EXTERNAL STANDARD	.095	19:02
BLANK TEST	.000	19:02
EXTERNAL STANDARD	.095	19:03
BLANK TEST	.000	19:03

N = 3
SIM. = .1
AVG. = .095

Operator Signature

 279

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
WASHINGTON POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 204004
10/29/13

ARREST TIME: 18:45

SUBJECT NAME:

N/A

DOB: 01/01/71 SEX: M

STATE/D.L.: MO/N/A

ARRESTING OFFICER:

N/A

OFFICER I.D.: 000

TESTING OFFICER:

KESTERSON/PAUL/R

OFFICER I.D.: 279

PERMIT NUMBER: 220340

EXPIRATION DATE: 10/10/14


MISCELLANEOUS DATA:

RFI CHECK

--- BREATH ANALYSIS ---

BLANK TEST	.000	19:06
INTERNAL STANDARD	VERIFIED	19:06
RADIO INTERFERENCE		

Operator Signature

 279

2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



PAUL R KESTERSON

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/10/2012

Number 220348

Expires 10/10/2014

MO 680-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)